

PATIENT INSTRUCTIONS

Knee Flexion Dynasplint® System Type II

Corporate Headquarters:
800.638.6771 toll-free
800.380.3784 fax

Canada:
800.668.9139 toll-free
905.851.3494 fax

Europe:
+31.45.523.0418 phone
+31.45.523.1303 fax

www.dynasplint.com

I acknowledge receipt of these patient instructions, which I have read and fully understand.

Signature

Print Name

Date

IMPORTANT: Read instructions thoroughly before wearing the Knee Flexion Dynasplint® System. Be sure that the splint fits comfortably and properly. If you feel pain, numbness, swelling, or skin irritation remove the splint immediately and contact your Dynasplint® Systems sales consultant. If joint stiffness persists longer than one hour after wearing the unit, reduce the tension by 0.5 increments, approximately 1/2 turn. Contact your Dynasplint® Systems sales consultant if you have any questions.

DYNASPLINT® and Dynasplint® Systems are registered trademarks of Dynasplint Systems, Inc.
Rev. 06/2017



Stretch Beyond Your Expectations.®


FITTING INSTRUCTIONS

APPLYING THE KNEE FLEXION DYNASPLINT® SYSTEM:

STEP 1: Completely open the shin and front-of-thigh cuffs.

STEP 2: Lay leg in the unit. The larger struts (with the windows and scales) should be placed on the lower leg with the counterforce (butterfly) strap in the crease of the knee. See Figure 1.


FIGURE 1.



STEP 3: Align the upper and lower struts with the midlines of the sides of the upper and lower leg. See Figure 1.

STEP 4: Align cams across the knee joint (joint axis). See Figure 2.

FIGURE 2.



STEP 5: Attach the Dynasplint System to the leg by feeding the front-of-thigh cuff through the D-wire on the opposite side. Pull back snugly; fasten the Velcro® sides together.

STEP 6: Repeat for the shin cuff.

STEP 7: Check cuffs for tightness. You should be able to slide a finger between the leg and the Dynasplint System cuffs and counterforce (butterfly) strap.

STEP 8: Check alignment. See Figures 1 and 2. The struts should line up with the centers of the sides of the leg, and the cams should be at the center of the sides of the knee.

STEP 9: Do not tamper with calf and back-of-thigh cuffs.

STEP 10: While the Dynasplint System is in place, you should be inactive and relaxed. The most comfortable position while sleeping is lying on the side with a pillow between the legs; when awake, sitting with a wedge pad under the thigh being sure not to block the lower leg.

REMOVING THE KNEE FLEXION DYNASPLINT® SYSTEM:

STEP 1: Completely open the shin and front-of-thigh cuffs.

STEP 2: Remove leg from unit.

Continue to follow your physician's or therapist's instructions regarding your home therapy program while wearing the Dynasplint System. Record your comments daily on the evaluation sheet (back page of booklet) for review with your Dynasplint Systems sales consultant.

CLINICIAN'S PRESCRIBED TREATMENT SCHEDULE

These are guidelines only. Remove the Dynasplint System if you experience pain at any time and contact your Dynasplint® Systems sales consultant.

Tension to be initially set at _____ increments.

Patient will wear the Dynasplint System for _____minutes/hours the first day.

If no more than one hour post-wear discomfort occurs, the patient may increase wear time according to the recommended daily wearing schedule, building up to an overnight application of 6-8 hours. Time spent wearing the splint is the most important component in regaining range of motion.

Increase tension by 0.5 – 1 increment on both sides of the splint if less than one hour of post-wear discomfort occurs after use.

Decrease tension if unable to wear for extended period of time.

Maximum tension: _____.

The basic protocol outline is to provide maximum benefit from the Dynasplint System. *Increasing tension faster does not ensure proper stretch will be applied.*

Remember to wear the Dynasplint System while inactive, preferably while sleeping. If unable to sleep in splint, maximize wear time during the day.

Your follow-up visit is _____.

Please bring the Dynasplint System and the completed evaluation sheet.

Sales Consultant: _____ Voicemail Number: _____

DYNASPLINT SYSTEMS PATIENT EVALUATION SHEET

Date	Hours of Wear	Tension Setting	Daily Goal	Duration of Post-Wear Stiffness	Comments

Continue to report your progress on a separate sheet of paper.

4 POINT SUMMARY

1. NO PAIN while wearing the Dynasplint System.
2. Time is the most important component.
3. Increase tension after 6-8 hours of wear if no more than one hour post-wear discomfort.
4. Decrease tension if unable to wear.

©1996 Dynasplint Systems, Inc.

Dynasplint Systems, Inc. products are covered by one or more of the following U.S.A. patents: 5,558,624; 5,645,521; 6,413,231; 6,506,172; 6,740,051; 6,908,475; 6,942,629; 6,942,631; 4,485,808; 4,508,111; 4,538,600; 4,944,290; 4,947,835; and 5,070,868. Other patents issued and/or pending in the U.S.A. and internationally.

The product described in the enclosed literature is intended for the specific purpose as per the instructions attached. Any use of this product outside of its intended purpose on any body part or in a manner outside the protocol established by Dynasplint Systems, Inc., is a use of the product for which it, its divisions and employees cannot be held responsible. All implied warranties of fitness for use for any other purpose (or purposes) are expressly disclaimed.